

PREGÃO ELETRÔNICO PARA REGISTRO DE PREÇOS Nº 13/2026 - ESCLARECIMENTOS md01

1 mensagem

Hexis Licitacoes <licitacoes@hexis.com.br>

28 de agosto de 2026 às 16:4

Para: "licitacaoaguasdopantanal@gmail.com" <licitacaoaguasdopantanal@gmail.com>

Prezados,

Em relação ao **Item 32**, gostaríamos de esclarecer um ponto técnico.

Ofertamos equipamentos que possui faixa de leitura de **0,02 a 2,00 mg/L** e **0,1 a 8,0 mg/L**, divergindo do edital apenas para medições **acima de 8,0 mg/L**.

Considerando que o equipamento atende integralmente às demais especificações técnicas e à aplicação pretendida, solicitamos confirmar se será aceito o fornecimento com a faixa de medição informada, em substituição à faixa requerida no edital.

Att,



MATEUS DANI | ANALISTA DE LICITAÇÕES

Hexis Científica | [E mdani@hexis.com.br](mailto:mdani@hexis.com.br)

T +55 (11) 4589-2614 | **SAC** 0800 702 2600

hexis.com.br | [Siga-nos no LinkedIn](#)

-



Como está nosso atendimento? Por favor, contate vozdocliente@hexis.com.br

Please be advised that this email may contain confidential information. If you are not the intended recipient, please notify us by email by replying to the sender and delete this message. The sender disclaims that the content of this email constitutes an offer to enter into, or the acceptance of, any agreement; provided that the foregoing does not invalidate the binding effect of any digital or other electronic reproduction of a manual signature that is included in any attachment.

Confidential - Company Proprietary